

# Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

CERTIFICATE OF STERILITY

Sterility Lot Number	Product Release Date:	Sterility Cert No.
CS20250219-US	21st March 2025	2500595

Dose Info	Min Specified	Max Specified	Min Actual	Max Actual
(KgY)	27.5	42	28.7	40.5

Process Code	M2500494/0495/0496/0497/0498/0499
--------------	-----------------------------------

ITEM NUMBER(S)	SIZE(S)
CTL414CS	XL
CTL417CS	SM, MD, LG
CTL428CS	2X
CTL903CSP	2X

Lakeland Industries Inc, doing business as "Lakeland Fire and Safety", certifies that the product/item number(s) and lot number(s) identified above have been processed and released in accordance with all relevant specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1 Sterilization of Health Care Products – Requirements for Validation and Routine Control – to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility are available online [www.lakeland.com](http://www.lakeland.com) or contact Customer Services at [info@lakeland.com](mailto:info@lakeland.com)

**D. Rumble, BSc, PCQI**

Global Compliance and Product Manager

